



**CERTIFICATE OF ANALYSIS**

StatSensor Creatinine Control Solution Level 3  
Lot Number 5023067243

Ref Number 43923  
Exp Date 2025-03-08

Date of Manufacture 2023-03-08

UDI# (01)00385480439237(11)230308(17)250308(10)5023067243

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	5.50 - 7.00	6.02

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Awishi M*

Approval

QC Inspector  
Title

5/19/2023  
Date