OI-922

Page 55 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

TFICATE OF COMPLIANCE Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:			NIPRO DIAGNOSTICS INC.
Description:			TEST STRIPS, NIPRO STATSTRIP TM
			GLUCOSE, PACKAGED
Catalog Number:			47940
Nova Biomedical P/	N		47940
Target Range (mg/dL)	Mean MPI	E (%) CV (%)
60-75	≤15 ≤	6.5	PASS / FAIL
120-160	≤15 ≤	5.5	PASS / FAIL
250-290	≤15 ≤	5.0	PASS / FAIL
380-420	≤ 15 ≤	5.0	PASS / FAIL
500-540	≤15 ≤	5.0	PASS / FAIL
Combined MPE (%)	≤10%		PASS / FAIL
Qty:			303
Lot No.			2123 30309
Mfg. Date:		3	2023.05.08
Expiration Date:			2025.11.30
MA			

Quality Control Manager

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454

OI-922

Page 56 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2 **RELEASE TEST RESULTS**

Customer:	NIPRO DIAGNOSTICS INC.	
Description:	TEST STRIPS, NIPRO STATSTRIP TM GLUCOSE	
Catalog Number:	47940	
Nova Biomedical P/N	47940	
Lot No.	2123130309	

	(Spiked whole blood)	Target Range:60-75	
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≦6.5		≦15
64	5.5	67-9	6.8

	(Spiked whole blood)	Target Range:120-160	6
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≦5.5		≦15
152	2-9	150-8	2-6

	(Spiked whole blood)	Target Range:250-290	
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≦5.0		≦15
284	3-3	281 -3	2-7

	(Spiked whole blood)	Target Range:380-420	
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≦5.0		≦15
407	3.0	414-0	2-6

	(Spiked whole blood)	Target Range:500-540	
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≦5.0		≦15
518	3.4	516-5	2-7

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454