

OI-922

Page 55 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

*nova*  
biomedical

**CERTIFICATE OF COMPLIANCE** Page 1/2

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE, PACKAGED
Catalog Number:	47940
Nova Biomedical P/N	47940
Target Range (mg/dL) Mean MPE (%) CV (%)	
60-75 ≤ 15 ≤ 6.5	PASS / FAIL
120-160 ≤ 15 ≤ 5.5	PASS / FAIL
250-290 ≤ 15 ≤ 5.0	PASS / FAIL
380-420 ≤ 15 ≤ 5.0	PASS / FAIL
500-540 ≤ 15 ≤ 5.0	PASS / FAIL
Combined MPE (%) ≤ 10%	PASS / FAIL
Qty:	304
Lot No.	2123131309
Mfg. Date:	2023.05.04
Expiration Date:	2025.11.30

Mac Broome  
Quality Control Manager

5/22/2023  
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454

OI-922 Page 56 of 66 Rev. BH Vialled and Packaged Test Strip Release Test Strips

*nova*  
biomedical

**ADDENDUM TO CERTIFICATE OF COMPLIANCE Page 2/2**

**RELEASE TEST RESULTS**

Customer:	NIPRO DIAGNOSTICS INC.
Description:	TEST STRIPS, NIPRO STATSTRIP™ GLUCOSE
Catalog Number:	47940
Nova Biomedical P/N	47940
Lot No.	2123131309


(Spiked whole blood) Target Range:60-75			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤6.5		≤15
71	3.5	69.8	3.0

(Spiked whole blood) Target Range:120-160			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.5		≤15
153	4.8	147.8	5.4

(Spiked whole blood) Target Range:250-290			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
284	4.6	280.5	3.7

(Spiked whole blood) Target Range:380-420			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
414	2.9	414.5	2.2

(Spiked whole blood) Target Range:500-540			
Mean (mg/dL) N=72	CV%	YSI (reference, mg/dL))	Mean MPE%
	≤5.0		≤15
524	3.1	523.3	2.4

  
 Quality Control Manager Date 5/22/2023

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454