



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

LifeScan Ref Number 02307701

Lot Number 0423135301

Exp Date 2025-11-15

Date of Manufacture 2023-05-15

UDI# (01)00385480417419 (11)230515(17) 251115(10)0423135301

	Spec	Mean Result
Glucose mg/dL	57 - 65	62

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A. Mishra

Approval

QC Inspector

Title

6/16/2023

Date