

CERTIFICATE OF ANALYSIS

Performance Check Solution Lot Number 23170035 Ref Number 17457 Exp Date 2024-12-19 Date of Manufacture 2023-06-19 UDI# (01)00385480174572(11)230619(17)241219(10)23170035

	<u>Spec</u>	Pass/Fail
Glu mg/dL	689.5 - 710.5	Pass
BUN mg/dL	97.0 - 103.0	Pass
Creatinine mg/dL	1.54 - 1.66	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi.M

Approval

QC Inspector

Title

6/26/2023

Date