



CERTIFICATE OF ANALYSIS

Nova Lactate Plus Control Solution Level 2
Lot Number 3823121302

Ref Number 40814
Exp Date 2025-11-01
Date of Manufacture 2023-05-01
UDI# (01)00385480408141(11)230501(17)251101(10)3823121302

| | <u>Spec</u> | <u>Mean Result</u> |
|----------------|-------------|--------------------|
| Lactate mmol/L | 4.4 - 5.0 | 4.7 |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

7/11/2023
Date