



**CERTIFICATE OF ANALYSIS**

StatSensor Creatinine Control Solution Level 1  
Lot Number 5023130241

Ref Number 43921

Exp Date 2025-05-10

Date of Manufacture 2023-05-10

UDI# (01)00385480439213(11)230510(17)250510(10)5023130241

	<u>Spec</u>	<u>Mean Result</u>
Creatinine mg/dL	0.90 - 1.10	1.02

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufactu

*Auishi M*

Approval

QC Inspector

Title

7/12/2023

Date