



**CERTIFICATE OF ANALYSIS**

Glucose pHOX plus Membrane Kit  
Lot Number 22033010

Part Number 33871  
Exp Date 2023-05-02  
Date of Manufacture 2022-02-02

|              | <u>Spec*</u> | <u>Mean Result</u> |
|--------------|--------------|--------------------|
| Slope        | 5.00 - 45.00 | 22.07              |
| pHOx Plus L3 | 79-93        | 84                 |
| Set A L1     | ≥470         | 512                |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

\*Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

2/19/2022  
Date