

## **CERTIFICATE OF ANALYSIS**

STP pHOx Plus L Cal Cartridge Lot Number 21348065

Ref Number 34088 Exp Date 2023-05-17 Date of Manufacture 2021-11-17 UDI# (01)00385480340885(11)211117(17)230517(10)21348065

Cal A L/N Na mmol/L HCO3 mmol/L K mmol/L Cl mmol/L Glu mg/dL Lac mmol/L	21324013 <u>Spec</u> 142 - 144 24.42 - 25.42 4.03 - 4.09 97.2 - 98.8 1.01 - 1.03 39.6 - 40.4 0.8 - 1.2	Pass/Fail Pass Pass Pass Pass Pass Pass Pass	Cal B L/N HCO3 mmol/L	21326044 <u>Spec</u> 45.90 - 47.78	<u>Pass/Fail</u> Pass	Cal C L/N pH @ 37C Ca mmol/L Na mmol/L K mmol/L CI mmol/L Glu mg/dL Lac mmol/L	21326045 <u>Spec</u> 7.332 - 7.338 2.01 - 2.07 72.5 - 73.5 10.04 - 10.20 52.6 - 53.4 198 - 202 5.5 - 6.5	Pass/Fail Pass Pass Pass Pass Pass Pass Pass
	-	Cal D L/N pH @ 37C Glu mg/dL .ac mmol/L	21326040 <u>Spec</u> 6.788 - 6.794 990 - 1010 10.5 - 13.5	<u>Pass/Fail</u> Pass Pass Pass	KCI	Reference L/N <u>Spec</u> 1.8 - 2.2	21321037 <u>Pass/Fail</u> Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

TAN

Approval

QC Inspector Title

<u>12/27/2021</u> Date