

OI-922

Page 30 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

| | |
|--|--|
| Customer: | N/A |
| Description: | TEST STRIPS, STATSTRIP A CREATININE, 50 COUNT |
| Catalog Number: | 58788 |
| Nova Biomedical P/N | 58788 |
| Control Level 1: CV (%) \leq 15% Control Level 2: CV (%) \leq 10% Control Level 3: CV (%) \leq 10% | PASS / FAIL PASS / FAIL PASS / FAIL |
| Qty: | 504 |
| Lot No. | 1323177129 |
| Mfg. Date: | 2023-06-26 |
| Expiration Date: | 2024-06-26 |
| Unique Device Identifier (UDI): | (01)20385480587888 (11)230626 (17)240626 (10)1323177129 |

Marc Brune
Quality Control Manager

7/20/2023
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454