

OI-922

Page 31 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

| | |
|--|--|
| Customer: | N/A |
| Description: | TEST STRIPS, STATSTRIP A CREATININE SELF-TEST, 50 COUNT |
| Catalog Number: | 60737 |
| Nova Biomedical P/N | 60737 |
| Control Level 1: CV (%) \leq 15% Control Level 2: CV (%) \leq 10% Control Level 3: CV (%) \leq 10% | <u>PASS</u> / FAIL <u>PASS</u> / FAIL <u>PASS</u> / FAIL |
| Qty: | 504 |
| Lot No. | 1323177129 |
| Mfg. Date: | 2023-06-26 |
| Expiration Date: | 2024-06-26 |
| Unique Device Identifier (UDI): | (01)20385480607371 (11) <u>2</u> <u>3</u> <u>0</u> <u>6</u> <u>2</u> <u>6</u> (17) <u>2</u> <u>4</u> <u>0</u> <u>6</u> <u>2</u> <u>6</u> (10) <u>1</u> <u>3</u> <u>2</u> <u>3</u> <u>1</u> <u>7</u> <u>7</u> <u>1</u> <u>2</u> <u>9</u> |

Mark Browne
Quality Control Manager

7/20/2023
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454