



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ pH Sensor
Lot Number 23142039

Part Number 48580
Exp Date 2024-11-22
Date of Manufacture 2023-05-22

UDI# (01)00385480485807(11)230522(17)241122(10)23142039

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.20 - 11.20	Pass
pHOx Ctl L1	7.178 - 7.228	Pass
pHOx Ctl L3	7.524 - 7.578	Pass
Std C mV	0.00 ± 110.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/25/2023
Date