

OI-922 Page 26 of 66 Rev. BH Vial and Packaged Test Strip Release Test Strips



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, STATSENSOR™ CREATININE, 50 COUNT
Catalog Number:	43370
Nova Biomedical P/N	43370
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS FAIL PASS FAIL PASS FAIL
Qty:	400
Lot No.	4923177129
Mfg. Date:	2023-06-26
Expiration Date:	2024-06-26
Unique Device Identifier (UDI):	(01)20385480433703 (11)230626 (17)240626 (10)4923177129


Quality Control Manager

7/25/2023
Date

Nova Biomedical Corporation 200 Prospect Street Waltham, MA 02454