



CERTIFICATE OF ANALYSIS

Bioprofile 300 Lactate Membrane
Lot Number 22167015

Part Number 34573
Exp Date 2023-06

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.0 - 60.0	46.9
Biopro Ctl L 1	0.89 - 1.39	1.09
Biopro Ctl L 2	2.55 - 3.45	2.94
Biopro Ctl L 3	4.25 - 5.75	4.59

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

Title

6/21/2022

Date