



CERTIFICATE OF ANALYSIS

Prime Reference Cartridge
Lot Number 23201046

Part Number 42043
Exp Date 2025-01-20
Date of Manufacture 2023-07-20
UDI# (01)00385480420433(11)230720(17)250120(10)23201046

	<u>Spec</u>	<u>Pass/Fail</u>
Na+ Slope	9.50 - 12.00	Pass
K+ Slope	9.00 - 11.50	Pass
Hct Slope	2.00 - 3.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

8/3/2023
Date