

OI-922

Page 26 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

*nova*  
biomedical

## CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	TEST STRIPS, STATSENSOR™ CREATININE, 50 COUNT
Catalog Number:	43370
Nova Biomedical P/N	43370
Control Level 1: CV (%) ≤ 15% Control Level 2: CV (%) ≤ 10% Control Level 3: CV (%) ≤ 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	20,016
Lot No.	4923198129
Mfg. Date:	2023-07-17
Expiration Date:	2025-07-17
Unique Device Identifier (UDI):	(01)20385480433703 (11)230717 (17)250717 (10)4923198129

Mar Browne  
Quality Control Manager

7/26/23  
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454