



CERTIFICATE OF ANALYSIS
Nova StatStrip Glucose and β -Ketone Control Solution Level 2

Nova Reference 46948
LifeScan Reference 02312901
Lot Number 0423163302

Exp Date 2025-12-12
Date of Manufacture 2023-06-12
UDI# (01)00385480469487(11)230612(17)251212(10)0423163302

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	105 - 125	111
Ketone mmol/L	0.7 - 1.1	0.8

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

08/17/2023
Date