OI-922

Page 65 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, STATSTRIP LACTATE, PACKAGED
Catalog Number:	47486
Nova Biomedical P/N	47486
Target Range CV (%) SD (mmol/L)	
$0.8-2.5 \text{ mmol/L} \leq 6.5\% \pm 0.3$	PASS FAIL
$3.0-5.0 \text{ mmol/L} \leq 5.5\% \qquad \pm 0.5$	PASS FAIL
10-14 mmol/L $\leq 5.0\%$ ± 1.2	PASS FAIL
Qty:	9,846
Lot No.	2623207249
Mfg. Date:	2023 - 07 - 26
Expiration Date:	2025 - 07 - 26
Unique Device Identifier (UDI):	(01)20385480 <u>47 4867</u> (11) <u>23 07 26</u> (17) <u>25 07 26</u> (10) 26 23 20 7 249

Quality Control Manager

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454