



CERTIFICATE OF ANALYSIS

Bioprofile Flex Linearity Solution HIGH
Lot Number 23201050

Catalog Number 45279
Exp Date 2025-01

| | <u>Spec</u> | <u>Pass/Fail</u> |
|------------|---------------|------------------|
| Gln mmol/L | 4.60 - 5.40 | Pass |
| Glu mmol/L | 4.60 - 5.40 | Pass |
| Gluc g/L | 13.50 - 16.50 | Pass |
| Lac g/L | 4.60 - 5.40 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Gluc, Lac

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

8/15/2023

Date