



CERTIFICATE OF ANALYSIS

BioPro Flex System Fluid Cartridge
Lot Number 22214066

Part Number 39580
Exp Date 2023-09

	S Sol L/N	22214065	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.900 - 7.100	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

8/5/2022

Date