



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Glucose Control Solution Level 2

Ref Number 41742

Lot Number 0423193302

Exp Date 2026-01-12

Date of Manufacture 2023-07-12

UDI# (01)00385480417426(11)230712(17)260112(10)0423193302

	Spec	Mean Result
Glucose mg/dL	105-125	115

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Auishi M*

Approval

QC Inspector

Title

8/30/2023

Date