



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 23240002

Part Number 48554
Exp Date 2024-11-28
Date of Manufacture 2023-08-28
UDI# (01)00385480485548(11)230828(17)241128(10)23240002

	<u>Spec*</u>	<u>Mean Result</u>
Slope 2	≥ 15.00	69.41
pHOx Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHOx Ultra/CCX Ctl L5	6.3 - 7.7	7.2
Lin Soln Set A Level 1	14.0 - 20.0	18.5

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

9/7/2023
Date