









5

Assay Data Sheet - Expected Ranges
BioProfile® FLEX Controls - LEVEL 5

5

Lot Number:  315195		Exp. Date: 2024-11	
Expected Range			
Parameter	Units	Lower Limit	Upper Limit
Gln	mmol/L	 4.41	 5.91
Glu	mmol/L	 1.78	 2.38
Gluc	g/L	 7.99	 11.99
Lac	g/L	 4.48	 5.98

nova
biomedical

200 Prospect Street • Waltham, MA U.S.A. 02454-9141 • (781) 864-0800