



CERTIFICATE OF ANALYSIS

Nova 13 Calibrator Pack
Lot Number 324115

Ref Number 15651
Exp Date 2025-08-29
Date of Manufacture 2023-08-29
UDI# (01)00385480156516(11)230829(10)324115

Cal A L/N 23241015 Cal B L/N 23241015 Reference L/N 23241015

	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	139.0 - 141.0	Pass	Na mmol/L	59.4 - 60.6	Pass	KCl mol/L	1.8 - 2.2	Pass
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass			
Cl mmol/L	113.6 - 115.4	Pass	Cl mmol/L	45.5 - 46.5	Pass			
Li mmol/L	0.49 - 0.51	Pass	Li mmol/L	4.95 - 5.05	Pass			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02453 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi M

Approval

QC Inspector
Title

9/8/2023
Date