



CERTIFICATE OF ANALYSIS

BioPro Flex Cell Density/Viability SA Reagent Pack
Lot Number 23255081

Ref Number 56260
Exp Date 2024-08

W Soln L/N 23250039

	<u>Spec</u>	<u>Result</u>
Depro	N/A	Present

C Soln L/N 23248058

	<u>Spec</u>	<u>Result</u>
3% Detergent	N/A	Present
pH	> 9.0	Pass

D Soln L/N 23248059

	<u>Spec</u>	<u>Result</u>
Depro	N/A	Present
%NaOCl	0.43 - 0.73	Pass

T Soln L/N 23249049

	<u>Spec</u>	<u>Result</u>
0.4% Trypan	N/A	Present
Particle Count	≤ 2.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/14/2023

Date