



CERTIFICATE OF ANALYSIS

NCRT 10 Calibrator Pack
Lot Number 320839

Ref Number 14608
Exp Date 2025-01-27
Date of Manufacture 2023-07-27
UDI# (01)00385480146081(11)230727(17)250127(10)320839

Cal A L/N	23208039		Cal B L/N	23208039	
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	139.0 - 141.0	Pass	Na mmol/L	59.4 - 60.6	Pass
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass
Cl mmol/L	113.6 - 115.4	Pass	Cl mmol/L	45.5 - 46.5	Pass
TCa mmol/L	1.94 - 2.06	Pass	TCa mmol/L	3.88 - 4.12	Pass
Li mmol/L	0.50 - 0.51	Pass	Li mmol/L	4.95 - 5.05	Pass

Ref L/N	23208039	
	<u>Spec</u>	<u>Result</u>
Ca++ Releasing Agent	N/A	Present

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi M

Approval

QC Inspector
Title

9/14/2023
Date