



CERTIFICATE OF ANALYSIS

Stat Profile pHox Plus L Calibrator Cartridge C
Lot Number 23264027

Ref Number 34087
Exp Date 2025-03-11

Date of Manufacture 2023-09-11

UDI# (01)00385480340878(11)230911(17)250311(10)23264027

Table with 3 columns: Cal-A L/N 23254093, Cal-B L/N 23254098, Cal-C L/N 23254099. Rows include analytes like Na mmol/L, HCO3 mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Glu mg/dL, Lac mmol/L with their respective specifications and pass/fail results.

Table for Cal-D L/N 23255010 with columns Spec and Pass/Fail. Rows include pH @ 37C, Glu mg/dL, Lac mmol/L.

Table for Reference L/N 23254094 with columns Spec and Pass/Fail. Row includes KCl mol/L.

This certifies that this product was manufactured and tested at Nova Biomedical-Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical-Devices Quality Management Systems Requirements, Medical-Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

9/25/2023
Date