



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ pO2 Sensor
Lot Number 23249074

Ref Number 48598
Exp Date 2025-09-06
Date of Manufacture 2023-09-06
UDI# (01)00385480485982(11)230906(17)250906(10)23249074

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------|----------------|------------------|
| Slope | -3.00 to -6.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

9/27/2023
Date