



**CERTIFICATE OF ANALYSIS**

Stat Profile® pHox Ultra™ / CCX™ Creatinine Membrane Kit  
Lot Number 23250008

Part Number 48597

Exp Date 2024-12-07

Date of manufacture 2023-09-07

UDI# (01)00385480485975(11)230907(17)241207(10)23250008

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 5.00	14.79
CtL 4	0.7 - 1.3	1.0
CtL 5	5.2 - 7.2	6.3

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

9/28/2023

Date