



**CERTIFICATE OF ANALYSIS**

Nova 8 Ionized Magnesium Sensor  
Lot Number 23241049

Ref Number 15265  
Exp Date 2024-11-29

Date of Manufacture 2023-08-29

UDI# (01)10385480152652(11)230829(17)241129(10)23241049

	<u>Spec*</u>	<u>Mean</u>
Slope	10.50 - 13.50	12.17
Chem Control L1 mmol/L	0.40 - 0.60	0.48
Set A Level 2 mmol/L	0.78 - 1.02	0.88

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

09/28/2023

Date