



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1
Lot Number 0422153301

Ref Number 41741
Exp Date 2024-12-02
Date of Manufacture 2022-06-02
UDI (01)00385480417419(11)220602(17)241202(10)0422153301

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector
Title

7/5/2022
Date