



**CERTIFICATE OF ANALYSIS**

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0422207301

LifeScan Ref Number 02307701

Exp Date 2025-01-26

Date of Manufacture 2022-07-26

	Spec	Mean Result
Glucose mg/dL	57-65	63

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

8/25/2022  
Date