



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1
Lot Number 0422259301

Ref Number 41741
Exp Date 2025-03-16
Date of Manufacture 2022-09-16
UDI (01)00385480417419(11)220916(17)250316(10)0422259301

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	62

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

10/13/2022
Date