



CERTIFICATE OF ANALYSIS

Acetate Membrane Kit
Lot Number 23271006

Part Number 33274
Exp Date 09/30/2025

	<u>Spec*</u>	<u>Mean Result</u>
Slope	7.5 - 16.0	9.1
Biopro Ctl L6	13.9 - 25.1	15.8
Biopro Ctl L7	75.0 - 125.0	108.1

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/6/2023
Date