



CERTIFICATE OF ANALYSIS

Stat Profile® pHox Ultra™ / CCX™ SO2 Sensor
Lot Number 23268024

Part Number 48591
Exp Date 2026-09-25
Date of manufacture 2023-09-25
UDI#(01)00385480485913(11)230925(17)260925(10)23268024

	<u>Spec</u>	<u>Mean Result</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	9.82	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

10/10/2023
Date