



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0423249301

LifeScan Ref Number 02307701

Exp Date 2026-03-06

Date of Manufacture 2023-09-06

UDI# (01)00385480417419 (11)230906(17)260306(10)0423249301

	Spec	Mean Result
Glucose mg/dL	57 - 65	63

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

10/5/2023

Date