



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 500 Sample  
Lot Number 23275025

Ref Number 53464  
Exp Date 2025-03-15  
Date of Manufacture 2023-09-15  
UDI# (01)00385480534642(11)230915(17)250315(10)23275025

Cal A	L/N	23262063	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.356	N/A
pCO2 mmHg		24.9	N/A

Cal B	L/N	23262064	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	23262077	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		51.6	N/A
pO2 mmHg		97.3	N/A

Reference	L/N	23258043	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

10/12/2023  
Date