



CERTIFICATE OF ANALYSIS

Nova 8 Ionized Magnesium Sensor
Lot Number 23256033

Ref Number 15265
Exp Date 2024-12-13

Date of Manufacture 2023-09-13

UDI# (01)10385480152652(11)230913(17)241213(10)23256033

	<u>Spec*</u>	<u>Mean</u>
Slope	10.5 - 13.5	12.4
Chem Control L1 mmol/L	0.40 - 0.60	0.47
Set A Level 2 mmol/L	0.78 - 1.02	0.88

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

10/12/2023

Date