OI-922 Page 33 of 66

ge 33 of 66 Rev. BH

Vialed and Packaged Test Strip Release

Test Strips



CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NA
Description:	TEST STRIPS, STATSTRIP TM
•	GLUCOSE, PACKAGED 1800 COUNT
Catalog Number:	42214
9	
Nova Biomedical P/N	42214
Target Range (mg/dL) Mean MPE (%) CV (%)	
60-75 ≤ 15 ≤ 6.5	PASS / FAIL
100.150	Disco / DAV
$120-160 \leq 15 \leq 5.5$	PASS / FAIL
$250-290 \leq 15 \leq 5.0$	PASS / FAIL
$380-420 \leq 15 \leq 5.0$	PASS / FAIL
500-540 ≤ 15 ≤ 5.0	PASS / FAIL
Combined MPE (%) ≤ 10%	PASS / FAIL
Qty:	- 0110
	3,842
Lot No.	0323228249
	1
Mfg. Date:	2023-08-16
Expiration Date:	2025-08-16
•	
Unique Device Identifier (UDI):	(01)20385480422141
*	(11) 2 3 0 8 / 6
	$(17) \frac{2}{5} \frac{5}{0} \frac{8}{8} \frac{7}{16}$
	(10) 0 3 2 3 2 2 8 2 4 9
A tacknowner	10/13/23

Quality Control Manager

Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454