



CERTIFICATE OF ANALYSIS

Nova 4/5/12/13/14 Chloride Sensor
Lot Number 23276013

Ref Number 14062
Exp Date 2025-01-03
Date of Manufacture 2023-10-03
UDI# (01)10385480140628(11)231003(17)250103(10)23276013

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	10.2 - 12.5	Pass
Nova Ctl L 1 mmol/L	103.0 - 109.0	Pass
Set A L 3 mmol/L	110.4 - 118.4	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

²Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

10/13/2023
Date