



**CERTIFICATE OF ANALYSIS**

StatStrip Glucose Control Level 3

Nova Reference 41743  
LifeScan Reference 02315601  
Lot Number 0422287303

Date of Manufacture 2025-04-14  
Exp Date 2022-10-14

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	300

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

11/28/2022  
Date