



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ Chloride Sensor  
Lot Number 23256032

Part Number 48557  
Exp Date 2024-12-13

Date of Manufacture 2023-09-13

UDI# (01)10385480485576(11)230913(17)241213(10)23256032

|             | <u>Spec</u>   | <u>Pass/Fail</u> |
|-------------|---------------|------------------|
| Slope       | 8.50 - 13.00  | Pass             |
| pHOx Ctl L4 | 121.3 - 130.3 | Pass             |
| pHOx Ctl L5 | 91.5 - 100.5  | Pass             |
| Std C mV    | 10.00 - 90.00 | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

10/20/2023  
Date