



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Level 3

Nova Reference 41743
LifeScan Reference 02315601
Lot Number 0423026303

Date of Manufacture 2023-01-26
Exp Date 2025-07-26

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	302

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

2/16/2023
Date