



**CERTIFICATE OF ANALYSIS**

STP Prime ABG Cal Cartridge 200 Sample  
Lot Number 23293013

Ref Number 53360  
Exp Date 2025-04-10  
Date of Manufacture 2023-10-10  
UDI# (01)00385480533607(11)231010(17)250410(10)23293013

Cal A	L/N	23283051		
		<u>Assigned Value*</u>	<u>Result</u>	
pH		7.354	N/A	
pCO2 mmHg		24.8	N/A	

Cal B	L/N	23283052		
		<u>Spec</u>	<u>Pass/Fail</u>	
pH		6.837 - 6.843	Pass	
pO2 mmHg		167.0	N/A	

Flush	L/N	23283040		
		<u>Assigned Value*</u>	<u>Result</u>	
pCO2 mmHg		51.6	N/A	
pO2 mmHg		97.4	N/A	

Reference	L/N	23284026		
		<u>Spec</u>	<u>Pass/Fail</u>	
KCl		1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Haneen Aldarwish*

Approval

QC Inspector

Title

10/27/2023

Date