



CERTIFICATE OF ANALYSIS

BioProfile Control Level 9
Lot Number 23237007

Catalog Number 43473
Exp Date 2024-11

	Spec	Pass/Fail
Particles/mL	55.00 x 10 ⁵ - 65.00 x 10 ⁵	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector
Title

10/31/2023
Date