



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0423276301

LifeScan Ref Number 02307701

Exp Date 2026-04-03

Date of Manufacture 2023-10-03

UDI # (01)00385480417419(11)231003(17)260403(10)0423276301

	Spec	Mean Result
Glucose mg/dL	57 - 65	60

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector

Title

10/31/2023

Date