



CERTIFICATE OF ANALYSIS

Linearity Std Set A Level 6
Lot Number 23298037

Ref Number 14071
Exp Date 2025-04-25
Date of Manufacture 2023-10-25
UDI# (01)00385480140713(11)231025(17)250425(10)23298037

	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	98.0 - 102.0	Pass
K+ mmol/L	98.0 - 102.0	Pass
Cl- mmol/L	196.0 - 204.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

11/1/2023

Date