



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ pCO2 Sensor
Lot Number 23291036

Ref Number 48600
Exp Date 2025-10-18
Date of Manufacture 2023-10-18
UDI# (01)00385480486002(11)231018(17)251018(10)23291036

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----------|---------------|------------------|
| Slope | 9.40 - 11.60 | Pass |
| Std C mV | 0.00 ± 150.00 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

11/6/2023
Date