



**CERTIFICATE OF ANALYSIS**

Allegro PT / INR Test Cartridge  
Lot Number 23306019

Ref Number 58006  
Exp Date 2024-11-02  
Date of Manufacture 2023-11-02

UDI# (01)00385480580069(11)231102(17)241102(10)23306019

Ctl Level 1			Ctl Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
PT sec.	7.5 - 14.5	Pass	PT sec.	17.0 - 35.0	Pass
INR	0.70 - 1.50	Pass	INR	1.60 - 3.60	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

*Haneen Aldarwish*

Approval

QC Inspector

Title

11/9/2023

Date