



**CERTIFICATE OF ANALYSIS**

STP® pHox Ultra™ / CCX™ Chloride Sensor  
Lot Number 23293047

Part Number 48557  
Exp Date 2025-01-20

Date of Manufacture 2023-10-20

UDI# (01)10385480485576(11)231020(17)250120(10)23293047

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	8.50 - 13.00	Pass
pHOx Ctl L4	121.3 - 130.3	Pass
pHOx Ctl L5	91.5 - 100.5	Pass
Std C mV	10.00 - 90.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

11/13/2023

Date